

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

NOV 09 2009

Supervisor/Operator Performing the Verification Procedure:	
Name <u>GARY J. TELLER</u>	ID# <u>2281</u> Date <u>11-5-09</u>
A Agency <u>AST- D Detachment</u> Phone # <u>451-5100</u>	
Instrument Location <u>1979 Payer Rd. F6Ks. 99701</u>	
B Alco S/N <u>[REDACTED]</u> Target Value <u>.083</u> High Pressure <u>300</u>	
Alco Test Value Average <u>.074</u> <u>.075</u>	
1 st Alco 2 nd Alco	
Signature <u>[Signature]</u>	
(OVER)	

(Do Not write in the area below)

- I, Nita J. Bolz, after being first duly sworn, depose and state as follows.
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
 - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
 - (3) I am the Scientific Director of the State Breath Alcohol Program.
 - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
 - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
 - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 21st day of Dec, 2009.

[Signature]
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N

130125 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

Breath Test Record

Serial Number 130125 Model DataMaster
Location FBK0 AST/Fairbanks

Subject Information

Last Name Verificationofcal O.L. # N
First Name IN

Operator Information

Operator Name Tellep Gary J Case N
Operator # 2281 Department FBK0

Test Results

Alco Target .083 Alco Serial X301591
Nominal 0.079 Barometer 28.52 Test Type V
Dry Gas 1 0.074 Time: 07:33:14 Test Date 11/05/2009
Subject Alcohol 0.000 Time: 07:34:34 Test Time 07:31:34
Dry Gas 2 0.075 Time: 07:35:54 Report Y

Diagnostic Test Record

Serial Number 130125 Model DataMaster
Location FBK0 AST/Fairbanks

Diagnostic Test Results

Test Date 11/05/2009 Computer Okay
Test Time 07:38:03 Program Okay
Software Date 02/20/2001
Sample Chamber (C) 49
Breath Tube (C) 42
Barometer (in) 28.52
Flow Detector Okay
High Speed Pump Okay
Detector Okay
Filter Okay
Quartz Okay
Calibration Okay

gmb
11/14/09 ✓